
Midmark M11 - v1.0.0
Total Cycles: 2646

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 177.0 F 2.5
4:00 185.2 F 6.4
6:00 213.6 F 12.0
8:00 224.6 F 7.0
10:00 237.6 F 12.5
12:00 236.5 F 10.0
14:00 238.8 F 8.1
16:00 250.1 F 15.6
18:00 262.8 F 22.9
20:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.5 F 29.1
2:00 272.1 F 29.1
3:00 272.4 F 29.0
4:00 271.9 F 29.0

Min 271.6 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 206.8 F 0.0
10:00 217.9 F 0.0
15:00 225.0 F 0.0
20:00 227.7 F 0.0
25:00 228.6 F 0.0
Dry Cycle Aborted

FILLING: 1:44
HEATING: 20:05
STERILIZING: 4:00
VENTING: 3:12
DRYING: 29:13
TOTAL CYCLE: 00:58:15

12/06/2022 12:06 PM
ACCEPTED: X
REJECTED:
Operator: U02