
Midmark M11 - v1.0.5
Total Cycles: 2603

Name: RITTER2
Sterilizer ID: V2313228
07/04/2023 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.8 F 0.0
2:00 145.3 F 0.5
4:00 162.4 F 3.3
6:00 185.5 F 7.2
8:00 208.8 F 12.3
10:00 220.6 F 5.6
12:00 231.4 F 10.0
14:00 237.4 F 9.1
16:00 241.7 F 12.4
18:00 240.9 F 11.2
20:00 253.5 F 17.6
22:00 264.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.6 F 29.5
2:00 271.6 F 28.9
3:00 272.7 F 29.6
4:00 271.8 F 28.9

Min 271.5 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 203.7 F 0.0
10:00 209.8 F 0.0
15:00 218.3 F 0.0
20:00 223.6 F 0.0
25:00 219.4 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:59
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

07/04/2023 02:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____