

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1387

Name: RITTER2  
Sterilizer ID: V2313228  
08/09/2022 03:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
2:00 203.2 F 3.9  
4:00 214.1 F 10.0  
6:00 225.6 F 6.5  
8:00 239.8 F 12.4  
10:00 240.3 F 11.0  
12:00 239.8 F 10.4  
14:00 254.6 F 17.9  
16:00 266.4 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.6 F 29.1  
2:00 272.1 F 29.1  
3:00 272.4 F 28.9  
4:00 271.9 F 29.0  
  
Min 271.6 F 28.5  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.0 F 0.0  
5:00 231.5 F 0.0  
10:00 215.7 F 0.0  
15:00 213.7 F 0.0  
20:00 210.5 F 0.0  
25:00 207.1 F 0.0  
30:00 206.9 F 0.0  
  
Drying Complete

FILLING: 2:34  
HEATING: 17:32  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:57:11  
CYCLE COMPLETE

08/09/2022 04:00 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_