

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3412

Name: RITTER2  
Sterilizer ID: V2313228  
02/26/2024 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.1 F 0.0  
2:00 130.7 F 0.7  
4:00 150.8 F 3.2  
6:00 174.9 F 7.2  
8:00 199.5 F 12.0  
10:00 223.7 F 5.9  
12:00 234.4 F 10.7  
14:00 230.5 F 7.2  
16:00 243.9 F 13.1  
18:00 241.8 F 11.7  
20:00 254.8 F 18.3  
22:00 266.4 F 25.2  
24:00 273.0 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.3 F 29.2  
2:00 272.9 F 29.6  
3:00 272.3 F 29.8  
4:00 271.8 F 29.1

Min 271.6 F 29.0  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.8 F 0.0  
5:00 231.5 F 0.0  
10:00 217.7 F 0.0  
15:00 216.4 F 0.0  
20:00 206.1 F 0.0  
25:00 206.0 F 0.0  
30:00 194.7 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 24:16  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:02:32  
CYCLE COMPLETE

02/26/2024 02:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_