
Midmark M11 - v1.0.0
Total Cycles: 1788

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2022 09:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.1 F 0.0
2:00 178.7 F 3.5
4:00 193.9 F 7.8
6:00 213.1 F 13.4
8:00 228.5 F 7.8
10:00 238.6 F 13.2
12:00 239.2 F 10.7
14:00 236.9 F 9.0
16:00 251.6 F 16.1
18:00 263.5 F 23.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.3
1:00 272.5 F 28.8
2:00 272.3 F 29.0
3:00 271.6 F 28.2
4:00 272.1 F 28.8

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 223.7 F 0.0
10:00 229.0 F 0.0
15:00 220.4 F 0.0
20:00 220.9 F 0.0
25:00 223.6 F 0.0
30:00 224.9 F 0.0

Drying Complete

FILLING: 1:35
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

04/04/2022 10:52 AM
ACCEPTED: X
REJECTED:
Operator: U02