
Midmark M11 - v1.0.0
Total Cycles: 1137

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2021 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 149.2 F 1.6
4:00 165.9 F 4.5
6:00 190.5 F 8.4
8:00 207.7 F 13.2
10:00 220.4 F 7.7
12:00 232.6 F 12.7
14:00 236.0 F 9.5
16:00 236.0 F 9.4
18:00 246.4 F 13.6
20:00 257.9 F 19.9
22:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.1
1:00 272.3 F 29.3
2:00 273.0 F 29.5
3:00 272.3 F 29.6
4:00 271.3 F 29.1

Min 271.3 F 29.0
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.6 F 0.0
5:00 229.9 F 0.0
10:00 228.6 F 0.0
15:00 229.2 F 0.0
20:00 227.1 F 0.0
25:00 219.9 F 0.0
30:00 212.3 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 23:51
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

08/31/2021 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____