
Midmark M11 - v1.0.5
Total Cycles: 2080

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 04:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.8 F 0.0
2:00 183.6 F 3.1
4:00 196.7 F 7.8
6:00 214.9 F 13.4
8:00 227.8 F 9.1
10:00 235.3 F 9.5
12:00 244.0 F 13.3
14:00 244.6 F 12.8
16:00 258.1 F 20.0
18:00 269.8 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.4 F 29.3
2:00 271.4 F 29.0
3:00 272.5 F 29.3
4:00 271.4 F 28.7

Min 271.2 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 229.9 F 0.0
10:00 229.3 F 0.0
15:00 224.3 F 0.0
20:00 221.9 F 0.0
25:00 224.3 F 0.0
30:00 226.0 F 0.0

Drying Complete

FILLING: 1:39
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:46
CYCLE COMPLETE

02/22/2023 05:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____