
Midmark M11 - v1.0.0
Total Cycles: 4219

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2024 04:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 70.2 F 0.3
4:00 86.0 F 1.2
6:00 123.8 F 3.3
8:00 144.7 F 6.0
10:00 165.9 F 9.4
12:00 187.3 F 13.1
14:00 203.5 F 4.3
16:00 215.0 F 7.7
18:00 225.7 F 11.4
20:00 227.9 F 6.2
22:00 239.7 F 11.4
24:00 236.4 F 8.6
26:00 249.9 F 15.0
28:00 260.7 F 21.2
30:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 272.6 F 28.7
4:00 271.8 F 28.5

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 206.4 F 0.0
10:00 211.5 F 0.0
15:00 217.3 F 0.0
20:00 217.8 F 0.0
25:00 218.3 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

01/23/2024 05:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____