
Midmark M11 - v1.0.0
Total Cycles: 4317

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2024 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.7 F 0.0
2:00 105.1 F 1.0
4:00 119.0 F 3.2
6:00 148.0 F 5.9
8:00 191.5 F 8.8
10:00 200.1 F 12.7
12:00 215.7 F 4.4
14:00 223.0 F 7.8
16:00 230.7 F 11.6
18:00 231.4 F 7.1
20:00 240.8 F 12.1
22:00 237.8 F 9.7
24:00 250.0 F 15.3
26:00 260.3 F 21.2
28:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.7 F 29.4
2:00 271.6 F 28.8
3:00 272.7 F 29.5
4:00 271.9 F 28.9

Min 271.6 F 28.8
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 209.8 F 0.0
10:00 216.1 F 0.0
15:00 218.6 F 0.0
20:00 217.0 F 0.0
25:00 215.9 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 29:45
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

02/21/2024 04:45 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____