
Midmark M11 - v1.0.5
Total Cycles: 3158

Name: RITTER2
Sterilizer ID: V2313228
11/23/2023 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.3 F 0.0
2:00 95.9 F 0.1
4:00 105.1 F 1.0
6:00 136.9 F 3.5
8:00 164.2 F 6.5
10:00 185.4 F 10.0
12:00 207.1 F 11.6
14:00 217.2 F 7.8
16:00 227.5 F 12.0
18:00 230.1 F 8.5
20:00 239.5 F 12.4
22:00 241.4 F 12.0
24:00 252.8 F 17.9
26:00 263.2 F 24.1
28:00 272.6 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.3
1:00 272.0 F 30.2
2:00 272.4 F 30.4
3:00 272.6 F 30.6
4:00 272.2 F 30.8

Min 271.6 F 30.2
Max 272.7 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 212.1 F 0.0
10:00 231.4 F 0.0
15:00 231.8 F 0.0
20:00 229.4 F 0.0
25:00 228.9 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:58
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:28
CYCLE COMPLETE

11/23/2023 01:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____