
Midmark M11 - v1.0.0
Total Cycles: 2264

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2022 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.3 F 0.0
2:00 198.0 F 1.8
4:00 204.0 F 6.8
6:00 221.2 F 12.6
8:00 229.2 F 8.0
10:00 241.7 F 13.0
12:00 241.9 F 11.8
14:00 241.1 F 10.9
16:00 255.4 F 18.1
18:00 267.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.5 F 28.8
2:00 271.6 F 28.4
3:00 272.5 F 28.8
4:00 271.6 F 28.1

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 223.2 F 0.0
10:00 232.0 F 0.0
15:00 229.1 F 0.0
20:00 225.8 F 0.0
25:00 219.0 F 0.0
30:00 205.8 F 0.0
Drying Complete

FILLING: 2:04
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

08/30/2022 10:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____