
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001497
14:59 10/13/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:39
PRESSURIZING 6:18
STERILIZING 7:43
136.0 C 220kPa 7:43
136.0 C 219kPa 8:13
136.5 C 223kPa 8:43
136.5 C 223kPa 9:13
136.4 C 223kPa 9:43
136.5 C 220kPa 10:13
136.4 C 223kPa 10:43
136.5 C 223kPa 11:13
VENTING 11:13
AIR DRYING 12:20
CYCLE COMPLETE 23:33

Digital Signature #
71A480D3EE42D3FF
