
Midmark M11 - v1.0.5
Total Cycles: 1570

Name: RITTER2
Sterilizer ID: V2313228
10/05/2022 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.7 F 0.0
2:00 189.2 F 0.1
4:00 200.9 F 3.1
6:00 220.7 F 8.3
8:00 233.9 F 11.6
10:00 237.3 F 10.7
12:00 236.1 F 9.0
14:00 239.4 F 9.8
16:00 249.7 F 15.4
18:00 262.1 F 22.5
20:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.2
2:00 271.6 F 29.0
3:00 272.6 F 29.3
4:00 271.6 F 28.7

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 222.5 F 0.0
10:00 250.5 F 0.0
15:00 227.2 F 0.0
20:00 223.0 F 0.0
25:00 216.6 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

10/05/2022 05:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____