
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001886
7:30 25/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:36
STERILIZING 3:43
136.0 C 220kPa 3:43
136.4 C 224kPa 4:13
136.4 C 221kPa 4:43
136.1 C 222kPa 5:13
136.4 C 224kPa 5:43
136.3 C 223kPa 6:13
136.0 C 221kPa 6:43
136.4 C 224kPa 7:13
136.4 C 224kPa 7:13
VENTING 7:13
AIR DRYING 8:08
CYCLE COMPLETE 8:13

Digital Signature #
2621921601758869
