
Midmark M11 - v1.0.0
Total Cycles: 1560

Name: RITTER 1
Sterilizer ID: V2178198
01/25/2022 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 189.0 F 3.6
4:00 187.1 F 8.2
6:00 206.8 F 13.2
8:00 220.0 F 7.7
10:00 234.3 F 12.9
12:00 237.0 F 10.2
14:00 236.6 F 8.1
16:00 249.5 F 15.2
18:00 261.6 F 22.0
20:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.8 F 29.6
2:00 272.4 F 29.0
3:00 272.2 F 29.3
4:00 272.8 F 29.6

Min 271.7 F 28.8
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 232.1 F 0.0
10:00 229.3 F 0.0
15:00 226.3 F 0.0
20:00 229.9 F 0.0
25:00 231.0 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

01/25/2022 01:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____