
Midmark M11 - v1.0.0
Total Cycles: 1481

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2021 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.7 F 0.0
2:00 210.2 F 3.5
4:00 207.7 F 8.3
6:00 221.7 F 9.2
8:00 230.7 F 9.4
10:00 238.4 F 9.2
12:00 243.0 F 12.6
14:00 241.8 F 11.2
16:00 254.8 F 18.0
18:00 266.3 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.6 F 29.2
2:00 271.8 F 28.4
3:00 272.1 F 29.2
4:00 272.5 F 29.0

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.5 F 0.0
5:00 223.7 F 0.0
10:00 236.9 F 0.0
15:00 229.9 F 0.0
20:00 228.0 F 0.0
25:00 227.3 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

12/21/2021 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____