

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1254

Name: RITTER2  
Sterilizer ID: V2313228  
07/05/2022 08:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.9 F 0.0  
2:00 118.7 F 0.5  
4:00 140.2 F 2.3  
6:00 157.8 F 5.8  
8:00 173.2 F 10.0  
10:00 217.3 F 5.8  
12:00 215.8 F 6.3  
14:00 226.0 F 10.4  
16:00 233.3 F 8.4  
18:00 238.9 F 11.7  
20:00 237.1 F 9.6  
22:00 249.2 F 15.7  
24:00 260.1 F 21.9  
26:00 270.3 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.8  
1:00 272.3 F 29.7  
2:00 272.7 F 30.0  
3:00 272.6 F 30.3  
4:00 272.0 F 29.8

Min 271.8 F 29.6  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.7 F 0.0  
5:00 194.7 F 0.0  
10:00 199.9 F 0.0  
15:00 214.0 F 0.0  
20:00 221.3 F 0.0  
25:00 222.5 F 0.0  
30:00 223.1 F 0.0  
Drying Complete

FILLING: 1:09  
HEATING: 27:38  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:05:59  
CYCLE COMPLETE

07/05/2022 09:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_