

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3440

Name: RITTER2  
Sterilizer ID: V2313228  
03/07/2024 01:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.3 F 0.0  
2:00 149.6 F 1.4  
4:00 158.3 F 4.2  
6:00 182.7 F 7.8  
8:00 199.2 F 11.9  
10:00 216.1 F 5.3  
12:00 226.0 F 9.6  
14:00 233.2 F 10.4  
16:00 239.4 F 11.8  
18:00 239.0 F 10.5  
20:00 251.5 F 16.7  
22:00 262.4 F 23.1  
24:00 272.5 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.2 F 29.7  
2:00 271.2 F 29.1  
3:00 272.2 F 29.9  
4:00 271.8 F 29.5

Min 271.0 F 29.1  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 210.9 F 0.0  
10:00 215.9 F 0.0  
15:00 209.9 F 0.0  
20:00 203.1 F 0.0  
25:00 200.1 F 0.0  
30:00 196.3 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:26  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:03:11  
CYCLE COMPLETE

03/07/2024 02:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_