
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000467
13:12 02/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:44
STERILIZING 4:16
136.4 C 233kPa 4:16
136.3 C 232kPa 4:46
136.2 C 231kPa 5:16
136.4 C 233kPa 5:46
136.1 C 231kPa 6:16
136.4 C 233kPa 6:46
136.3 C 233kPa 7:16
136.5 C 233kPa 7:46
136.5 C 233kPa 7:46
VENTING 7:46
AIR DRYING 8:42
CYCLE COMPLETE 8:54

Digital Signature #
1E11ABA94CBEE677
