
Midmark M11 - v1.0.5
Total Cycles: 3559

Name: RITTER2
Sterilizer ID: V2313228
04/25/2024 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.7 F 0.0
2:00 149.4 F 1.6
4:00 149.7 F 4.5
6:00 175.4 F 8.5
8:00 200.6 F 13.4
10:00 216.9 F 6.5
12:00 231.6 F 11.4
14:00 232.1 F 8.2
16:00 242.1 F 11.8
18:00 244.5 F 12.8
20:00 256.9 F 19.3
22:00 268.0 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.6 F 29.5
2:00 271.8 F 28.8
3:00 271.5 F 29.1
4:00 272.6 F 29.4

Min 271.1 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 227.0 F 0.0
10:00 199.6 F 0.0
15:00 193.5 F 0.0
20:00 190.0 F 0.0
25:00 188.2 F 0.0
30:00 186.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

04/25/2024 10:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____