
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001172
15:43 18/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:35
STERILIZING 4:47
136.0 C 221kPa 4:47
136.3 C 224kPa 5:17
136.2 C 222kPa 5:47
136.4 C 225kPa 6:17
136.4 C 224kPa 6:47
136.2 C 223kPa 7:17
136.4 C 225kPa 7:47
136.1 C 222kPa 8:17
136.1 C 222kPa 8:17
VENTING 8:17
AIR DRYING 9:10
CYCLE COMPLETE 9:13

Digital Signature #
76FCBB85539150B8
