
Midmark M11 - v1.0.0
Total Cycles: 2967

Name: RITTER 1
Sterilizer ID: V2178198
03/02/2023 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.2 F 0.0
2:00 169.7 F 0.1
4:00 192.3 F 0.4
6:00 199.6 F 4.3
8:00 216.2 F 8.8
10:00 229.1 F 12.3
12:00 234.0 F 9.7
14:00 236.1 F 8.9
16:00 244.7 F 12.5
18:00 245.5 F 13.0
20:00 258.5 F 20.1
22:00 270.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.6 F 29.0
2:00 271.6 F 28.4
3:00 272.8 F 29.2
4:00 272.1 F 28.6

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 213.6 F 0.0
10:00 235.6 F 0.0
15:00 236.3 F 0.0
20:00 226.5 F 0.0
25:00 217.4 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:52
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

03/02/2023 01:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____