

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4796

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2024 12:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.6 F 0.0  
2:00 159.3 F 0.9  
4:00 179.5 F 4.1  
6:00 218.3 F 7.8  
8:00 216.2 F 12.0  
10:00 223.4 F 5.2  
12:00 234.7 F 9.4  
14:00 239.3 F 11.0  
16:00 240.9 F 11.0  
18:00 237.0 F 8.9  
20:00 250.2 F 15.1  
22:00 260.8 F 21.0  
24:00 270.0 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 271.9 F 28.7  
2:00 272.5 F 28.9  
3:00 272.4 F 28.7  
4:00 271.9 F 28.4

Min 271.2 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.7 F 0.0  
5:00 225.8 F 0.0  
10:00 223.7 F 0.0  
15:00 226.4 F 0.0  
20:00 217.4 F 0.0  
25:00 173.5 F 0.0  
30:00 170.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:07  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:03:59  
CYCLE COMPLETE

08/08/2024 01:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_