

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000769  
10:15 09/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:41  
PRESSURIZING 3:42  
STERILIZING 4:51  
136.0 C 220kPa 4:51  
136.0 C 221kPa 5:21  
135.9 C 221kPa 5:51  
136.2 C 223kPa 6:21  
136.3 C 223kPa 6:51  
136.3 C 223kPa 7:21  
136.0 C 221kPa 7:51  
136.2 C 223kPa 8:21  
136.2 C 223kPa 8:21  
VENTING 8:22  
AIR DRYING 9:17  
CYCLE COMPLETE 9:19

-----  
Digital Signature #  
5A57F00F50C33B6C  
-----