
Midmark M11 - v1.0.0
Total Cycles: 2881

Name: RITTER 1
Sterilizer ID: V2178198
02/09/2023 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.0 F 0.0
2:00 188.5 F 4.2
4:00 192.0 F 9.0
6:00 215.6 F 7.9
8:00 227.0 F 9.0
10:00 234.4 F 9.8
12:00 241.2 F 12.0
14:00 240.5 F 10.7
16:00 254.3 F 17.6
18:00 266.1 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.3
2:00 272.3 F 28.9
3:00 271.5 F 28.3
4:00 271.6 F 28.7

Min 271.2 F 28.2
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 223.2 F 0.0
10:00 236.1 F 0.0
15:00 235.5 F 0.0
20:00 229.9 F 0.0
25:00 226.7 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 19:41
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

02/09/2023 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____