
Midmark M11 - v1.0.0
Total Cycles: 2730

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2023 02:38 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 178.5 F 2.4
4:00 187.6 F 7.7
6:00 208.4 F 10.0
8:00 228.5 F 8.8
10:00 234.1 F 8.3
12:00 242.9 F 11.5
14:00 246.9 F 13.5
16:00 261.1 F 21.3
18:00 272.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 271.8 F 28.0
2:00 272.5 F 28.4
3:00 272.8 F 28.8

Min 271.4 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 227.8 F 0.0
Dry Cycle Aborted

FILLING: 0:51
HEATING: 18:14
STERILIZING: 3:00
VENTING: 2:49
DRYING: 1:04
TOTAL CYCLE: 00:25:57

01/03/2023 03:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____