
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005271
10:59 27/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:29
STERILIZING 3:32
136.5 C 220kPa 3:32
136.1 C 221kPa 4:02
136.3 C 221kPa 4:32
136.4 C 224kPa 5:02
136.3 C 225kPa 5:32
136.4 C 222kPa 6:02
136.2 C 220kPa 6:32
136.4 C 221kPa 7:02
136.4 C 221kPa 7:02
VENTING 7:03
AIR DRYING 8:01
CYCLE COMPLETE 13:44

Digital Signature #
6D279D0F863134C7
