

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2302

Name: RITTER 1  
Sterilizer ID: V2178198  
09/08/2022 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 97.5 F 0.0  
2:00 117.8 F 0.4  
4:00 122.2 F 1.4  
6:00 128.9 F 4.4  
8:00 159.1 F 8.1  
10:00 186.2 F 12.4  
12:00 208.2 F 4.6  
14:00 220.8 F 8.4  
16:00 232.0 F 12.4  
18:00 232.4 F 8.4  
20:00 242.5 F 12.1  
22:00 242.3 F 11.6  
24:00 254.3 F 17.8  
26:00 265.2 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.6 F 29.3  
2:00 271.8 F 28.6  
3:00 271.9 F 29.2  
4:00 272.5 F 29.0

Min 271.2 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 197.7 F 0.0  
10:00 213.9 F 0.0  
15:00 220.4 F 0.0  
20:00 221.4 F 0.0  
25:00 227.6 F 0.0  
30:00 222.2 F 0.0

Drying Complete

FILLING: 0:58  
HEATING: 27:53  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:06:05  
CYCLE COMPLETE

09/08/2022 10:13 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_