
Midmark M11 - v1.0.0
Total Cycles: 1975

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2022 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.8 F 0.0
2:00 198.3 F 4.6
4:00 211.6 F 9.5
6:00 219.3 F 7.2
8:00 230.8 F 10.9
10:00 231.5 F 7.7
12:00 241.1 F 11.6
14:00 243.5 F 12.5
16:00 256.2 F 19.0
18:00 267.2 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.1 F 29.5
2:00 273.0 F 29.8
3:00 272.0 F 29.0
4:00 272.7 F 29.9

Min 271.0 F 28.6
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 220.1 F 0.0
10:00 239.8 F 0.0
15:00 241.0 F 0.0
20:00 234.6 F 0.0
25:00 213.0 F 0.0
30:00 186.5 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 19:32
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

05/30/2022 04:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____