
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002388
14:12 07/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:04
STERILIZING 4:10
136.5 C 225kPa 4:10
136.4 C 221kPa 4:41
136.4 C 224kPa 5:11
136.1 C 218kPa 5:41
136.4 C 223kPa 6:11
136.2 C 221kPa 6:41
136.4 C 223kPa 7:11
136.4 C 223kPa 7:41
136.4 C 223kPa 7:41
VENTING 7:41
AIR DRYING 8:32
CYCLE COMPLETE 8:46

Digital Signature #
BCB8DF66433D789E
