
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005157
14:56 27/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:01
STERILIZING 4:06
136.5 C 220kPa 4:06
136.4 C 221kPa 4:36
136.4 C 223kPa 5:06
136.2 C 224kPa 5:36
136.4 C 225kPa 6:06
136.2 C 220kPa 6:36
136.4 C 223kPa 7:06
136.3 C 224kPa 7:36
136.3 C 224kPa 7:36
VENTING 7:36
AIR DRYING 8:26
CYCLE COMPLETE 11:33

Digital Signature #
2C020DB412CEDC15
