
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002124
11:19 04/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:29
STERILIZING 3:17
136.0 C 220kPa 3:17
136.4 C 226kPa 3:47
136.2 C 224kPa 4:17
136.4 C 225kPa 4:47
136.4 C 225kPa 5:17
136.4 C 226kPa 5:47
136.4 C 222kPa 6:17
136.3 C 225kPa 6:47
136.3 C 225kPa 6:47
VENTING 6:47
AIR DRYING 7:38
CYCLE COMPLETE 13:04

Digital Signature #
F11E1A405CFCDE99
