
Midmark M11 - v1.0.0
Total Cycles: 2261

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.5 F 0.0
2:00 194.5 F 0.9
4:00 214.3 F 6.2
6:00 230.9 F 12.3
8:00 235.0 F 9.3
10:00 236.7 F 8.8
12:00 242.4 F 11.0
14:00 248.0 F 14.0
16:00 261.3 F 21.4
18:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.4 F 28.4
2:00 272.0 F 28.5
3:00 272.3 F 28.3
4:00 272.1 F 28.5

Min 271.5 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 231.7 F 0.0
10:00 228.6 F 0.0
15:00 227.9 F 0.0
20:00 228.2 F 0.0
25:00 230.4 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:18
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:57:32
CYCLE COMPLETE

08/29/2022 04:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____