
Midmark M11 - v1.0.0
Total Cycles: 1387

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2021 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 184.4 F 3.4
4:00 206.9 F 8.9
6:00 227.2 F 8.0
8:00 235.6 F 11.1
10:00 235.6 F 8.8
12:00 238.4 F 9.4
14:00 248.8 F 14.7
16:00 262.0 F 22.0
18:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.2 F 28.9
2:00 271.3 F 28.4
3:00 272.4 F 28.9
4:00 271.3 F 28.2

Min 271.1 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.0 F 0.0
5:00 241.9 F 0.0
10:00 221.2 F 0.0
15:00 211.2 F 0.0
20:00 206.0 F 0.0
25:00 223.6 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 18:14
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

11/18/2021 03:51 PM
ACCEPTED: X
REJECTED:
Operator: U02