

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3963

Name: RITTER 1  
Sterilizer ID: V2178198  
11/01/2023 07:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 71.3 F 0.2  
4:00 98.1 F 1.1  
6:00 113.3 F 3.2  
8:00 140.3 F 5.8  
10:00 164.5 F 8.8  
12:00 183.1 F 12.2  
14:00 199.7 F 2.5  
16:00 211.7 F 5.7  
18:00 221.7 F 9.0  
20:00 231.2 F 12.8  
22:00 231.7 F 8.2  
24:00 242.7 F 13.0  
26:00 240.7 F 10.8  
28:00 252.6 F 16.7  
30:00 262.9 F 22.7  
32:00 272.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.1 F 28.9  
2:00 272.2 F 28.7  
3:00 271.5 F 28.5  
4:00 272.6 F 29.1  
  
Min 271.4 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.8 F 0.0  
5:00 200.2 F 0.0  
10:00 206.2 F 0.0  
15:00 206.3 F 0.0  
20:00 210.1 F 0.0  
25:00 213.4 F 0.0  
30:00 217.2 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:18  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:10:39  
CYCLE COMPLETE

11/01/2023 08:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_