

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2003

Name: RITTER2  
Sterilizer ID: V2313228  
02/02/2023 04:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.2 F 0.0  
2:00 193.1 F 4.3  
4:00 198.3 F 9.3  
6:00 218.3 F 7.3  
8:00 230.5 F 10.5  
10:00 231.8 F 8.1  
12:00 239.9 F 10.8  
14:00 246.8 F 14.4  
16:00 259.8 F 21.7  
18:00 271.6 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.4 F 29.8  
2:00 272.1 F 30.0  
3:00 272.1 F 29.6  
4:00 272.2 F 30.1  
  
Min 271.5 F 29.3  
Max 272.6 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.1 F 0.0  
5:00 217.1 F 0.0  
10:00 230.4 F 0.0  
15:00 229.2 F 0.0  
20:00 230.6 F 0.0  
25:00 220.9 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 1:59  
HEATING: 18:34  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

02/02/2023 05:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_