
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003349
7:40 31/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:33
STERILIZING 3:38
136.4 C 221kPa 3:38
136.1 C 222kPa 4:08
136.2 C 219kPa 4:38
136.4 C 221kPa 5:08
136.2 C 222kPa 5:38
136.0 C 221kPa 6:08
135.7 C 219kPa 6:38
136.2 C 219kPa 7:08
136.2 C 219kPa 7:08
VENTING 7:08
AIR DRYING 8:00
CYCLE COMPLETE 8:05

Digital Signature #
FEBF5E55503BD9B6
