
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 008224
14:24 22/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:20
STERILIZING 4:28
136.4 C 222kPa 4:28
136.4 C 223kPa 4:58
136.3 C 223kPa 5:28
136.2 C 219kPa 5:58
136.3 C 222kPa 6:28
136.3 C 224kPa 6:58
136.4 C 224kPa 7:28
136.1 C 220kPa 7:58
136.1 C 220kPa 7:58
VENTING 7:58
AIR DRYING 8:51
CYCLE COMPLETE 38:51

Digital Signature #
9781FDB10D514C6A
