
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003921
15:26 29/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:56
STERILIZING 2:50
136.0 C 220kPa 2:50
136.4 C 224kPa 3:20
136.3 C 223kPa 3:50
136.0 C 222kPa 4:20
136.4 C 222kPa 4:50
136.4 C 224kPa 5:20
136.4 C 223kPa 5:50
136.3 C 224kPa 6:20
136.3 C 224kPa 6:20
VENTING 6:20
AIR DRYING 8:24
CYCLE COMPLETE 11:00

Digital Signature #
9C0B20AB9FFE4E0C
