
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006331
10:56 20/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:53
PRESSURIZING 1:56
STERILIZING 2:54
136.0 C 220kPa 2:54
136.4 C 223kPa 3:25
136.3 C 222kPa 3:55
136.4 C 224kPa 4:25
136.4 C 223kPa 4:55
136.4 C 220kPa 5:25
136.4 C 224kPa 5:55
136.0 C 219kPa 6:25
136.0 C 219kPa 6:25
VENTING 6:25
AIR DRYING 7:14
CYCLE COMPLETE 7:18

Digital Signature #
A744A491C2526911
