

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007322  
8:51 06/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:46  
PRESSURIZING 3:47  
STERILIZING 4:47  
136.0 C 220kPa 4:47  
136.0 C 219kPa 5:17  
136.3 C 224kPa 5:47  
136.3 C 224kPa 6:17  
136.2 C 223kPa 6:47  
136.2 C 221kPa 7:17  
136.3 C 223kPa 7:47  
136.2 C 222kPa 8:17  
136.2 C 222kPa 8:17  
VENTING 8:17  
AIR DRYING 9:12  
CYCLE COMPLETE 9:21

-----  
Digital Signature #  
C9C2E77E4918F5E3  
-----