
Midmark M11 - v1.0.5
Total Cycles: 3584

Name: RITTER2
Sterilizer ID: V2313228
05/02/2024 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 169.3 F 3.1
4:00 188.2 F 7.0
6:00 205.0 F 11.2
8:00 215.8 F 5.3
10:00 229.7 F 9.9
12:00 233.8 F 9.5
14:00 242.0 F 12.4
16:00 241.2 F 11.3
18:00 253.2 F 17.4
20:00 263.5 F 23.6
22:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.3 F 29.7
2:00 271.9 F 29.1
3:00 271.1 F 29.0
4:00 272.5 F 30.2

Min 271.0 F 28.8
Max 273.2 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 222.8 F 0.0
10:00 216.7 F 0.0
15:00 213.0 F 0.0
20:00 212.5 F 0.0
25:00 210.1 F 0.0
30:00 210.8 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:01:60
CYCLE COMPLETE

05/02/2024 03:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____