
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006162
12:49 18/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:47
STERILIZING 3:51
136.0 C 220kPa 3:51
136.4 C 222kPa 4:21
136.3 C 222kPa 4:51
136.4 C 221kPa 5:21
136.4 C 223kPa 5:51
136.3 C 222kPa 6:21
136.3 C 223kPa 6:51
136.0 C 222kPa 7:21
136.0 C 222kPa 7:21
VENTING 7:21
AIR DRYING 8:10
CYCLE COMPLETE 13:40

Digital Signature #
509D2F20BB366916
