
Midmark M11 - v1.0.0
Total Cycles: 3385

Name: RITTER 1
Sterilizer ID: V2178198
06/14/2023 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 181.6 F 0.3
4:00 187.0 F 4.6
6:00 211.3 F 10.0
8:00 224.3 F 6.4
10:00 231.3 F 10.7
12:00 231.4 F 7.5
14:00 243.1 F 12.2
16:00 244.3 F 12.6
18:00 256.5 F 19.0
20:00 267.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.7 F 29.2
2:00 271.8 F 28.6
3:00 271.3 F 28.6
4:00 272.7 F 29.2

Min 271.1 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.2 F 0.0
5:00 218.1 F 0.0
10:00 229.9 F 0.0
15:00 228.0 F 0.0
20:00 227.3 F 0.0
25:00 227.1 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 21:29
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

06/14/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____