

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2383

Name: RITTER 1  
Sterilizer ID: V2178198  
10/03/2022 07:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.5 F 0.0  
2:00 80.1 F 0.0  
4:00 93.4 F 0.0  
6:00 133.4 F 1.0  
8:00 156.8 F 2.8  
10:00 174.7 F 5.2  
12:00 192.0 F 8.2  
14:00 206.7 F 11.8  
16:00 217.1 F 4.4  
18:00 226.3 F 8.4  
20:00 236.4 F 12.7  
22:00 235.6 F 9.0  
24:00 241.0 F 10.4  
26:00 245.3 F 12.8  
28:00 256.9 F 19.0  
30:00 267.4 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.2 F 28.6  
2:00 272.4 F 29.0  
3:00 271.6 F 28.2  
4:00 272.4 F 28.8

Min 271.5 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.3 F 0.0  
5:00 218.9 F 0.0  
10:00 232.7 F 0.0  
15:00 230.6 F 0.0  
20:00 226.9 F 0.0  
25:00 229.5 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 31:26  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:09:23  
CYCLE COMPLETE

10/03/2022 08:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_