
Midmark M11 - v1.0.5
Total Cycles: 3370

Name: RITTER2
Sterilizer ID: V2313228
02/08/2024 02:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.8 F 0.0
2:00 94.4 F 0.0
4:00 107.8 F 0.7
6:00 131.6 F 2.9
8:00 158.6 F 6.0
10:00 177.6 F 9.5
12:00 197.2 F 13.3
14:00 215.5 F 6.6
16:00 229.0 F 10.7
18:00 229.1 F 7.6
20:00 238.7 F 12.1
22:00 238.6 F 10.4
24:00 250.6 F 16.4
26:00 261.5 F 22.7
28:00 271.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.6 F 29.9
2:00 271.2 F 29.2
3:00 272.2 F 30.0
4:00 271.8 F 29.5

Min 271.1 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 230.3 F 0.0
10:00 237.3 F 0.0
15:00 217.5 F 0.0
20:00 204.5 F 0.0
25:00 191.7 F 0.0
30:00 195.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:35
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:07:03
CYCLE COMPLETE

02/08/2024 03:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____