
Midmark M11 - v1.0.0
Total Cycles: 2832

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2023 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 168.1 F 3.4
4:00 176.7 F 7.5
6:00 197.7 F 12.4
8:00 217.2 F 5.9
10:00 228.8 F 10.8
12:00 232.2 F 7.1
14:00 244.2 F 13.2
16:00 242.9 F 11.5
18:00 255.6 F 18.2
20:00 267.0 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.0 F 28.7
2:00 271.6 F 28.7
3:00 271.8 F 28.5
4:00 271.7 F 28.8

Min 271.2 F 28.2
Max 272.3 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 231.0 F 0.0
10:00 231.8 F 0.0
15:00 232.8 F 0.0
20:00 231.6 F 0.0
25:00 228.5 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:46
CYCLE COMPLETE

01/31/2023 12:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____