
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000719
10:25 21/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:15
STERILIZING 4:20
136.5 C 232kPa 4:20
136.4 C 233kPa 4:50
136.4 C 230kPa 5:20
136.3 C 232kPa 5:50
136.4 C 234kPa 6:20
136.1 C 230kPa 6:50
136.4 C 233kPa 7:20
136.4 C 233kPa 7:50
136.4 C 233kPa 7:50
VENTING 7:50
AIR DRYING 8:45
CYCLE COMPLETE 9:02

Digital Signature #
488D7DCB9BD13924
