
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001722
13:23 09/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:58
STERILIZING 3:59
136.4 C 221kPa 3:59
136.1 C 223kPa 4:29
136.3 C 224kPa 4:59
136.0 C 222kPa 5:29
136.5 C 226kPa 5:59
136.2 C 220kPa 6:29
136.2 C 219kPa 6:59
136.4 C 226kPa 7:29
136.4 C 226kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 8:24

Digital Signature #
7D4BFA1570C1CBF0
