
Midmark M11 - v1.0.5
Total Cycles: 3397

Name: RITTER2
Sterilizer ID: V2313228
02/21/2024 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.0 F 0.0
2:00 125.9 F 0.3
4:00 137.5 F 2.5
6:00 169.1 F 6.0
8:00 184.8 F 10.2
10:00 213.5 F 7.3
12:00 224.1 F 8.6
14:00 232.5 F 13.3
16:00 235.3 F 10.4
18:00 235.1 F 8.5
20:00 248.1 F 15.0
22:00 259.5 F 21.3
24:00 269.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.8
1:00 272.4 F 29.8
2:00 272.5 F 30.1
3:00 272.4 F 30.3
4:00 271.9 F 30.1

Min 271.7 F 29.7
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 222.0 F 0.0
10:00 233.4 F 0.0
15:00 214.7 F 0.0
20:00 203.1 F 0.0
25:00 196.6 F 0.0
30:00 193.5 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:38
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:04:21
CYCLE COMPLETE

02/21/2024 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____